

Tables

Table1 Stimulation Level Calibration Results of Different Signals in OR and SPC

Signal	S1	S2	S3
Duration	1 ms	2 ms	3 ms
Signal pattern	0.5-0-0.5	1-0-1	1.5-0-1.5
Average DRD	-0.02±0.19	0.01±0.17	0.02±0.14
Correlation coefficientin SPC	0.995 (P<.001)	0.996 (P<.001)	0.998 (P<.001)
Correlation coefficient in OR	0.999 (P<.001)	0.998 (P<.001)	0.999 (P<.001)

Table1:

OR: operating room; SPC: sound-proof chamber; DRD: dial to reference difference